

Terminalsteuerung im FPII_3: Kurzanleitung

Main Menu

- A **Display status** →
- B **Display/Modify parameters** →
- C Format
- E Streamer tubes conditioning
- G HV Modules MAP
- I Crate # [0]
- L Set protections
- N Clear alarm

Also valid: 1=Top, 2=Back

Display Status

	V _{mon}	I _{mon}	V ₀	I ₀	
Block					
Oben					
Mitte					
Unten					

Also valid: 1=Top, 2=Back

V₀ block [2700]:

V₀ oben [1000]:

V₀ Mitte [1000]:

V₀ unten [1000]:

Spannung ändern:

neue Spannung-RETURN

Spannung beibehalten:

RETURN

Display/Modify Parameters

- A Single Channel
- C **Group** →
- E Global absolute
- F Global relative

Also valid: 1=Top, 2=Back

Display/Modify Parameters: Group

- A Group name
- C **V₀** →
- D V₁
- F I₀
- G I₁
- I Ramp_up
- J Ramp_down
- L Trip
- N **Status** →
- Q Next group

Also valid: 1=Top, 2=Back

Status "block" is off
Turn it on? [No]

⋮
Status "oben" is off
Turn it on? [No]

⋮
Status "Mitte" is off
Turn it on? [No]

⋮